

## **observation volume, $V_{\text{obs}}$**

*in flame emission and absorption spectrometry*

The volume of that part of the flame that is observed through the optical device (in  $\mu\text{l}$ ).

**See also:** observation space

### **Source:**

PAC, 1986, 58, 1737 (*Quantities and units in clinical chemistry: Nebulizer and flame properties in flame emission and absorption spectrometry (Recommendations 1986)*) on page 1741