

## **geometry factor**

*in radioanalytical chemistry*

The average solid angle in steradians at source subtended by the aperture or sensitive volume of the detector, divided by  $4\pi$ .

**Source:**

PAC, 1994, 66, 2513 (*Nomenclature for radioanalytical chemistry (IUPAC Recommendations 1994)*) on page 2520