

## **value of a division**

*of a precision balance scale*

The reciprocal of the sensitivity; like the latter, it usually varies with load. It is determined by empirical calibration.

**Source:**

Orange Book, p. 35