

## **zero point**

*of a glass electrode*

Value of the pH of a solution, which in combination with a stated outer reference electrode, gives zero emf from the operational cell.

**Source:**

PAC, 1985, 57, 531 (*Definition of pH scales, standard reference values, measurement of pH and related terminology (Recommendations 1984)*) on page 541